

CALIBRATION CERTIFICATE

Customer Details-

Quick Skip Recycling Ltd
18 Thorn Business Park
Rotherwas
Hereford
HR2 6JT

Cal Cert Ref-	83519027-05/10/2026
---------------	---------------------

Scale Details		As Found Linearity & Hysteresis		
Location	Yard	LOAD	READING	DEVIATION
Serial Number	83519027	0kg	0kg	0
Manufacturer	Dini Argeo 3590ET	4000kg	4000kg	0
Capacity	50000kg x 20kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	05/10/2026	16000kg	16000kg	0
Next Calibration Due	10/2027	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	38420kg	38420kg	0
Issue date	05/10/2026	43300kg	43300kg	0

REPEATABILITY				
LOAD		READING 1	READING 2	READING 3
POSITION		READING 4	READING 5	READING 6
1	6880kg	6880kg	6880kg	6880kg
2	6880kg	6880kg	6880kg	6880kg
3	6880kg	6880kg	6880kg	6880kg
4	6880kg	6880kg	6880kg	6880kg
5	6880kg	6880kg	6880kg	6880kg
6	6880kg	6880kg	6880kg	6880kg

Compliance Weights Used		Calibration of above machine	
Box Set No	Cal Cert No	Tested By-	Jay Hutton
C098 - C117	TR13440	Signature-	
All test weights are calibrated and traceable to National Standards. Make up weights are also used as required			