

CALIBRATION CERTIFICATE

Customer Details-

Quick Skip Recycling Ltd
18 Thorn Business Park
Rotherwas
Hereford
HR2 6JT

Cal Cert Ref-

83519027-13/10/2025

Scale Details		As Found Linearity & Hysteresis		
Location	Yard	LOAD	READING	DEVIATION
Serial Number	83519027	0kg	0kg	0
Manufacturer	Dini Argeo 3590ET	4000kg	4000kg	0
Capacity	50000kg x 20kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	13/10/2025	16000kg	16000kg	0
Next Calibration Due	October 2026	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	30920kg	30920kg	0
Issue date	13/10/2025	43680kg	43680kg	0

REPEATABILITY

LOAD	READING 1	READING 2	READING 3	READING 4
43680kg	43680kg	43680kg	43680kg	43680kg

ECCENTRIC LOADING

POSITION	LOAD	READING	INDICATOR 1 2 3 4 5 6
1	6640kg	6640kg	
2	6640kg	6640kg	
3	6640kg	6640kg	
4	6640kg	6640kg	
5	6640kg	6640kg	
6	6640kg	6640kg	

COMPLIANCE WEIGHTS USED

Box Set No	Cal Cert No	Tested By:	Ashley Bryant
W25013 - (01 - 20)	W25013	Signature-	

All test weights are calibrated and traceable to National Standards. | Make up weights are also used as required