

CALIBRATION CERTIFICATE

Customer Details-

Quick Skip Recycling Ltd
18 Thorn Business Park
Rotherwas
Hereford
HR2 6JT

Cal Cert Ref-

83519027-31/10/2024

Scale Details		As Found Linearity & Hysteresis		
Location	Yard	LOAD	READING	DEVIATION
Serial Number	83519027	0kg	0kg	0
Manufacturer	Dini Argeo 3590ET	4000kg	4000kg	0
Capacity	50000kg x 20kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	31/10/2024	16000kg	16000kg	0
Next Calibration Due	October 2025	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	39720kg	39720kg	0
Issue date	31/10/2024	44380kg	44380kg	0

REPEATABILITY				
LOAD	READING 1	READING 2	READING 3	READING 4
44380kg	44380kg	44380kg	44380kg	44380kg

ECCENTRIC LOADING			
POSITION	LOAD	READING	INDICATOR
1	6660kg	6660kg	1 2 3 4 5 6
2	6660kg	6660kg	
3	6660kg	6660kg	
4	6660kg	6660kg	
5	6660kg	6660kg	
6	6660kg	6660kg	

Compliance Weights Used		Calibration of above machine	
Box Set No	Cal Cert No	Tested By-	Jay Hatton
W24017 - (01 - 20)	W24017 - Feb 24	Signature-	
All test weights are calibrated and traceable to National Standards. Make up weights are also used as required			