

# CALIBRATION CERTIFICATE

**Customer Details-**


**Quick Skip Recycling Ltd**  
**18 Thorn Business Park**  
**Rotherwas**  
**Hereford**  
**HR2 6JT**

|               |                              |
|---------------|------------------------------|
| Cal Cert Ref- | <b>83519027-verification</b> |
|---------------|------------------------------|

| Scale Details        |                     | As Found Linearity & Hysteresis |         |           |
|----------------------|---------------------|---------------------------------|---------|-----------|
| Location             | Yard                | LOAD                            | READING | DEVIATION |
| Serial Number        | <b>83519027</b>     | <b>0kg</b>                      | 0kg     | 0         |
| Manufacturer         | Dini Argeo 3590ET   | <b>4000kg</b>                   | 4000kg  | 0         |
| Capacity             | 50000kg x 20kg      | <b>8000kg</b>                   | 8000kg  | 0         |
| Machine Type         | Weighbridge         | <b>12000kg</b>                  | 12000kg | 0         |
| Calibration Date     | 29/10/2018          | <b>16000kg</b>                  | 16000kg | 0         |
| Next Calibration Due | <b>October 2019</b> | <b>20000kg</b>                  | 20000kg | 0         |
| Tolerance            | +/- 3 Divisions     | <b>29700kg</b>                  | 29700kg | 0         |
| Issue date           | 29/10/2018          | <b>43900kg</b>                  | 43900kg | 0         |

| REPEATABILITY  |           |           |           |           |
|----------------|-----------|-----------|-----------|-----------|
| LOAD           | READING 1 | READING 2 | READING 3 | READING 4 |
| <b>43900kg</b> | 43900kg   | 43900kg   | 43900kg   | 43900kg   |

| ECCENTRIC LOADING |                |         |           |   |
|-------------------|----------------|---------|-----------|---|
| POSITION          | LOAD           | READING | INDICATOR |   |
| 1                 | <b>10000kg</b> | 10000kg | 1         | 2 |
| 2                 | <b>10000kg</b> | 10000kg | 3         | 4 |
| 3                 | <b>10000kg</b> | 10000kg | 5         | 6 |
| 4                 | <b>10000kg</b> | 10000kg |           |   |
| 5                 | <b>10000kg</b> | 10000kg |           |   |
| 6                 | <b>10000kg</b> | 10000kg |           |   |

| Compliance Weights Used               |                         | Calibration of above machine |   |
|---------------------------------------|-------------------------|------------------------------|---|
| Box Set No                            | Cal Cert No             | Tested By-                   | Mark Foxall   |
| <b>01-22,47-50,1&amp;2, PM51-PM75</b> | <b>W17195,W18012-13</b> | Signature-                   |  |

All test weights are calibrated and traceable to National Standards.

Make up weights are also used  
 RIG – 19700kg FLT-4200kg