

CALIBRATION CERTIFICATE

Customer Details-


Quick Skip Recycling Ltd
18 Thorn Business Park
Rotherwas
Hereford
HR2 6JT

Cal Cert Ref-	83519027-06/11/2019
----------------------	---------------------

<i>Scale Details</i>		<i>As Found Linearity & Hysteresis</i>		
Location	Yard	LOAD	READING	DEVIATION
Serial Number	83519027	0kg	0kg	0
Manufacturer	Dini Argeo 3590ET	4000kg	4000kg	0
Capacity	50000kg x 20kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	06/11/2019	16000kg	16000kg	0
Next Calibration Due	November 2020	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	39940kg	39940kg	0
Issue date	06/11/2019	44180kg	44180kg	0

<i>REPEATABILITY</i>				
LOAD	READING 1	READING 2	READING 3	READING 4
44180kg	44180kg	44180kg	44180kg	44180kg

<i>ECCENTRIC LOADING</i>				
POSITION	LOAD	READING	INDICATOR	
1	6240kg	6240kg	1	2
2	6240kg	6240kg	3	4
3	6240kg	6240kg	5	6
4	6240kg	6240kg		
5	6240kg	6240kg		
6	6240kg	6240kg		

<i>Compliance Weights Used</i>		<i>Calibration of above machine</i>	
Box Set No	Cal Cert No	Tested By-	James Hudson
PM51-PM75, 47-50,1 & 2	W19001/W19002	Signature-	
All test weights are calibrated and traceable to National Standards.			
Make up weights are also used Rig 19940kg FLT 4240kg			