

CALIBRATION CERTIFICATE


Customer Details-
Quick Skip Recycling Ltd
18 Thorn Business Park
Rotherwas
Hereford
HR2 6JT

Cal Cert Ref-	83519027-02/10/2023
----------------------	---------------------

<i>Scale Details</i>		<i>As Found Linearity & Hysteresis</i>		
Location	Yard	LOAD	READING	DEVIATION
Serial Number	83519027	0kg	0kg	0
Manufacturer	Dini Argeo 3590ET	4000kg	4000kg	0
Capacity	50000kg x 20kg	8000kg	8000kg	0
Machine Type	Weighbridge	12000kg	12000kg	0
Calibration Date	02/10/2023	16000kg	16000kg	0
Next Calibration Due	October 2024	20000kg	20000kg	0
Tolerance	+/- 3 Divisions	39280kg	39280kg	0
Issue date	02/10/2023	43940kg	43940kg	0

<i>REPEATABILITY</i>				
LOAD	READING 1	READING 2	READING 3	READING 4
43940kg	43940kg	43940kg	43940kg	43940kg

<i>ECCENTRIC LOADING</i>									
POSITION	LOAD	READING	INDICATOR						
1	6660kg	6660kg	<table border="1"> <tr><td>1</td><td>2</td></tr> <tr><td>3</td><td>4</td></tr> <tr><td>5</td><td>6</td></tr> </table>	1	2	3	4	5	6
1	2								
3	4								
5	6								
2	6660kg	6660kg							
3	6660kg	6660kg							
4	6660kg	6660kg							
5	6660kg	6660kg							
6	6660kg	6660kg							

<i>Compliance Weights Used</i>		<i>Calibration of above machine</i>	
Box Set No	Cal Cert No	Tested By-	Jay Hatton
W23017 & W23018 - 02/2024	W23017/W23018	Signature-	
All test weights are calibrated and traceable to National Standards. Make up weights are also used as required			